


# Product information

Information about other products is available at: [www.demeditec.com](http://www.demeditec.com)



## User's Manual

# Hazelnut ELISA

Enzyme Immunoassay for the quantitative determination of Hazelnut in food



DEHAZE01



96

Sensitivity (Hazelnut)	0.3 ppm
Recovery	83-101 %
Incubation Time	60 min

## 1. GENERAL INFORMATION

Hazelnut (*Corylus avellana*) belongs to the birch plants. With 13 % the fraction of proteins in hazelnuts is high. Many of these proteins are known for being allergenic, such as Cor a 9 and Cor a 11. Most of them are very heat resistant making them stable to different production processes. For this reason hazelnut represents an important food allergen. For hazelnut allergic persons hidden hazelnut allergens in food are a critical problem. Already very low amounts of hazelnuts can cause allergic reactions, which may lead to anaphylactic shock in severe cases. Because of this, hazelnut allergic persons must strictly avoid the consumption of hazelnuts or hazelnut containing food. Crosscontamination, mostly in consequence of the production process is often noticed. The chocolate production process is a representative example. This explains why in many cases the existence of hazelnut residues in foods cannot be excluded. For this reason sensitive detection systems for hazelnut residues in foodstuffs are required.

The **Hazelnut ELISA** represents a highly sensitive detection system and is particularly cap-able of the quantification of hazelnut residues in cookies, cereals, ice cream and chocolate.

## 2. PRINCIPLE OF THE TEST

The **Hazelnut** quantitative test is based on the principle of the enzyme linked immunosorbent assay. An antibody directed against hazelnut proteins is bound on the surface of a microtiter plate. Hazelnut containing samples or standards are given into the wells of the microtiter plate. After 20 minutes incubation at room temperature, the wells are washed with diluted washing solution to remove unbound material. A peroxidase conjugated second antibody directed against hazelnut proteins is given into the wells and after 20 minutes of incubation the plate is washed again. A substrate solution is added and incubated for 20 minutes, resulting in the development of a blue colour. The colour development is inhibited by the addition of a stop solution, and the colour turns yellow. The yellow colour is measured photometrically at 450 nm. The concentration of hazelnut is directly proportional to the colour intensity of the test sample.

## 3. PRECAUTIONS

Full compliance of the following good laboratory practices (GLP) will determine the reliability of the results:

1. Prior to beginning the assay procedure, bring all reagents to room temperature (20-25°C).
2. All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
3. Once the assay has been started, all subsequent steps should be completed without interruption and within the recommended time limits.
4. Replace caps in all the reagents immediately after use. Do not interchange vial stoppers.
5. Use a separate disposable tip for each specimen to prevent cross-contamination.
6. All specimens and standards should be run at the same time, so that all conditions of testing are the same.
7. Do not mix components from different batches.
8. Do not use reagents after expiration date.
9. Check both precision and accuracy of the laboratory equipment used during the procedure (micropipets, ELISA reader etc.).

## 4. HEALTH AND SAFETY INSTRUCTIONS

1. Do not smoke or eat or drink or pipet by mouth in the laboratory.
2. Wear disposable gloves whenever handling patient specimens.
3. Avoid contact of substrate and stop solution with skin and mucosa (possible irritation, burn or toxicity hazard). In case of contact, rinse the affected zone with plenty of water.
4. Handling and disposal of chemical products must be done according to good laboratory practices (GLP).

## 5. REAGENTS

The kit contains reagents for 96 determinations. They have to be stored at 2-8°C. Expiry data are found on the labels of the bottles and the outer package.

1. Microtiter plate consisting of 12 strips with 8 breakable wells each, coated with anti-hazelnut antibodies.
2. Hazelnut Standards (0; 1; 4; 10; 40 ppm of hazelnut): 5 vials with 1.0 mL each, dyed red, ready-to-use
3. Conjugate (anti-hazelnut-peroxidase): 15 mL, dyed red, ready-to-use.
4. Substrate Solution (TMB): 15 mL, ready-to-use.
5. Stop Solution (1 N acidic solution): 15 mL, ready-to-use.
6. Extraction and sample dilution buffer (Tris): 2 x 120 mL as 10x concentrate, dyed red. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least one week. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
7. Washing Solution (PBS + Tween 20): 60 mL as 10x concentrate. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least 4 weeks. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
8. Plastic bag to store unused microtiter strips.
9. Instruction Manual.

## 6. ADDITIONAL INSTRUMENTATION AND REAGENTS (NOT PROVIDED)

### Instrumentation

- 100 - 1000 µL micropipets
- Volumetric flask
- Analytical balance
- Mortar, mixer
- Water bath
- Centrifuge
- ELISA reader (450 nm)

### Reagents

- double distilled water

## 7. SAMPLE PREPARATION

Due to high risk of cross-contamination all applied instruments like applicator, mortar, glass vials etc. have to be **cleaned thoroughly** before and after each sample. Hazelnut proteins adhere very strongly to different surfaces. In certain cases they can resist a common dishwasher cleaning. To identify possible cross-contamination caused by previous extractions it is strongly recommended to note the sequence of the extractions.

The following sample preparation should be applied for all kinds of samples:

1. To maximize homogeneity and representativeness of the sample drawing, a minimum of 5 g sample should be pulverized finely in a mortar, impact mill etc.
2. 1 g of the homogenized mixture is suspended in 20 mL of **pre-diluted** extraction buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes.
3. The samples are centrifuged for 10 minutes at 2000 g. If it is not possible to separate the supernatant from the precipitate completely, the suspension should be filtrated if necessary.
4. 100 µL of particle-free solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the **pre-diluted** extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

## 8. PROCEDURE

The washing solution is supplied as 10x concentrate and has to be **diluted** 1+9 with double distilled water before use.

In any case the **ready-to-use** standards provided should be determined twofold. When samples in great quantities are determined, the standards should be pipetted once before the samples and once after the samples. For final interpretation the arithmetic mean is used for calculation.

In consideration of GLP and quality control requirements a duplicate measurement of samples is recommended.

The procedure is according to the following scheme:

1. Prepare samples as described above.
2. Pipet 100 µL **ready-to-use** standards or prepared samples in duplicate into the appropriate wells of the microtiter plate.
3. Incubate for 20 minutes at room temperature.
4. Wash the plate three times as follows: Discard the contents of the wells (dump or aspirate). Pipet 300 µL of diluted washing solution into each well. After the third repetition empty the wells again and remove residual liquid by striking the plate against a paper towel. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbencies.
5. Pipet 100 µL of conjugate (anti-hazelnut-peroxidase) into each well.
6. Incubate for 20 minutes at room temperature.
7. Wash the plate as outlined in 4.
8. Pipet 100 µL of substrate solution into each well.
9. Allow the reaction to develop in the dark (e.g. cupboard or drawer; the chromogen is light-sensitive) for 20 minutes at room temperature.
10. Stop enzyme reaction by adding 100 µL of stop solution (1 N acidic solution) into each well. The blue colour will turn yellow upon addition.
11. After thorough mixing, measure absorbance at 450 nm (reference wavelength 620 nm), using an ELISA reader. The colour is stable for 30 minutes.

## 9. CALCULATION OF RESULTS

The ready-to-use standards are prepared for a direct determination of sample concentrations. The dilution of samples in the extraction process as described in the above stated sample preparation procedure is already considered. Additional dilution due to high sample concentration has to be accounted for.

1. Calculate the average optical density (OD 450 nm) for each set of reference standards or samples.
2. Construct a standard curve by plotting the mean optical density obtained for each reference standard against its concentration in ppm on semi-log graph paper with the optical density on the vertical (y) axis and the concentration on the horizontal (x) axis. Alternatively the evaluation can be carried out by software. In this case the 4-parameter method should be preferred.
3. Using the mean optical density value for each sample, determine the corresponding concentration of hazelnut in ppm from the standard curve. Depending on experience and/or the availability of computer capability, other methods of data reduction may be employed.

## 10. TYPICAL STANDARD VALUES

The following table contains an example for a typical standard curve. The binding is calculated as percent of the absorption of the 40 ppm standard. These values are only an example and should not be used instead of the standard curve which has to be measured in every new test.

Hazelnut (ppm)	% binding of 40 ppm
40	100
10	41
4	23
1	9
0	5

## 11. PERFORMANCE

### Sensitivity

The limit of detection (LOD) of the **Hazelnut test** is 0.3 ppm.

The limit of quantification (LOQ) of the **Hazelnut test** is 1 ppm.

Due to the variety of sample matrices and their influence on the blank, results less than the LOQ should be treated as negative.

### Cross-reactivity

For the following foods no cross-reactivity could be detected:

Wheat	Soy	Brazil nut
Barley	Poppy seed	Pistachio
Rye	Sunflower seed	Macadamia nut
Oats	Pumpkin seed	Chestnut
Buckwheat	Pine nuts	Cocoa
Corn	Cashew nut	Dried milk
Rice	Sesame	Gluten
Pea	Peanut	Lecithin
Chickpea	Walnut	Gelatin
Bean	Coconut	Apple
Almond		

### Precision

Intra-assay Precision	7 – 12 %
Inter-assay Precision	3 – 13 %

### Linearity

The serial dilution of spiked samples (cookies, cereals, ice cream and chocolate) resulted in a dilution linearity of 87% - 121%.

### Recovery

Mean recovery was determined by spiking samples with different amounts of hazelnut:

Cookies	101 %
Cereals	99 %
Ice cream	90 %
Chocolate	83 %

**12. REFERENCES**

1. Wigotzki M, et al. (2000) – Influence of varieties, storage and heat treatment on IgE-binding proteins in hazelnuts (*Corylus avellana*). *Food and Agric Immunol*, 12:217-29
2. Le TM, et al. (2008) – Reported food allergy to peanut, tree nuts and fruit: comparison of clinical manifestations, prescription of medication and impact on daily life. *Allergy*, 63(7):910-6
3. Luettkopf D, et al. (2001) – Comparison of four variants of major allergen in hazelnut (*Corylus avellana*) Cor a 1.04 with the major hazel pollen allergen Cor a 1.01. *Mol Immunol*, 38:515-25
4. Ballmer-Weber BK, et al. (2008) – Predictive value of the sulfidoleukotriene release assay in oral allergy syndrome to celery, hazelnut, and carrot. *J Investig Allergol Clin Immunol*, 18(2):93-99
5. Ben Rejeb S, et al. (2005) – Multi-allergen screening immunoassay for the detection of protein markers of peanut and four tree nuts in chocolate. *Food Addit Contam*, 22(8):709-15
6. Lauer I, et al. (2004) – Hazelnut (*Corylus avellana*) vicilin Cor a 11: molecular characterization of a glycoprotein and its allergenic activity. *Biochem J*, 383:327-334
7. Holzhauser T, et al. (2002) – Detection of potentially allergenic hazelnut (*Corylus avellana*) residues in food: a comparative study with DNA PCR-ELISA and protein sandwich-ELISA. *J Agric Food Chem*, 50(21):5808-15
8. Kiening M, et al. (2005) – Sandwich immunoassays for the determination of peanut and hazelnut traces in foods. *J Agric Food Chem*, 53(9):3321-7
9. Holzhauser T, et al. (1999) – Quantitative sandwich ELISA for determination of traces of hazelnut (*Corylus avellana*) protein in complex food matrixes. *J Agric Food Chem*, 47(10):4209-18
10. Akkerdaas JH, et al. (2004) – A novel approach for the detection of potentially hazardous pepsin stable hazelnut proteins as contaminants in chocolate-based food. *J Agric Food Chem*, 52(25):7726-31

Empfindlichkeit	0,3 ppm
Wiederfindung	83 – 101 %
Inkubationszeit	60 min

## 1. ALLGEMEINES

Die Haselnuss (*Corylus avellana*) gehört zu den Birkengewächsen und hat mit ca. 13% einen hohen Proteinanteil im Kern. Viele dieser Proteine sind als allergieauslösend bekannt, wie z.B. Cor a 9 und Cor a 11. Desweiteren sind viele dieser Proteine hitzeresistent, was sie stabil gegenüber verschiedenen Produktionsprozessen macht. Aus diesem Grund stellt die Haselnuss ein bedeutsames Nahrungsmittelallergen dar. Für Haselnuss-Allergiker sind versteckte Haselnussallergene in Nahrungsmitteln ein kritisches Problem. Schon sehr geringe Mengen von Haselnuss lösen allergische Reaktionen bis hin zum anaphylaktischen Schock aus. Aus diesem Grund müssen Haselnussallergiker auf den Konsum von Haselnüssen oder haselnusshaltigen Nahrungsmitteln strikt verzichten. Aufgrund von Kreuzkontaminationen, meist bedingt durch den Produktionsprozess von Nahrungsmitteln wie z.B. bei Schokolade, kann bei einigen Lebensmitteln das Vorhandensein von Haselnussrückständen nicht ausgeschlossen werden. Um diese detektieren zu können, bedarf es sensitiver Nachweissysteme. Der **Haselnuss Test** stellt ein hochsensibles Nachweissystem dar und ist insbesondere zur Quantifizierung von Haselnussrückständen in Keksen, Cerealien, Eiscreme und Schokolade geeignet.

## 2. TESTPRINZIP

Der **Haselnuss Test** basiert auf dem Prinzip des Enzymimmunoassay (EIA). Ein gegen Haselnussprotein gerichteter Antikörper ist auf der Oberfläche der Mikrotiterplatte gebunden. Darauf wird die Haselnuss enthaltende Probe bzw. Standards in die Vertiefungen der Mikrotiterplatte gegeben. Es findet eine Bindung zwischen Antikörper und Haselnussprotein statt. Nach 20-minütiger Inkubation bei Raumtemperatur wird die Platte mit verdünntem Waschpuffer gewaschen, um nichtgebundenes Material zu entfernen. Ein Peroxidase-markierter, gegen Haselnussprotein gerichteter zweiter Antikörper wird in die Vertiefungen pipettiert. Nach einer weiteren 20-minütigen Inkubation wird erneut gewaschen. Eine Substratlösung wird hinzupipettiert und 20 Minuten inkubiert, wodurch in den Vertiefungen ein blauer Farbstoff entwickelt wird. Die Farbentwicklung wird durch Zugabe einer Stopp-Lösung beendet, wobei ein Farbumschlag von blau nach gelb stattfindet. Die resultierende Farbe wird spektrophotometrisch bei 450 nm gemessen. Die Haselnusskonzentration ist der Intensität der Färbung direkt proportional.

## 3. VORSICHTSMAßNAHMEN

Die volle Übereinstimmung mit den folgenden Regeln für eine gute Laborpraxis (GLP) wird zu vertrauenswürdigen Ergebnissen führen:

1. Vor der Testdurchführung müssen alle Reagenzien auf Raumtemperatur gebracht werden (20°C-25°C).
2. Alle Reagenzien sollten vor der Verwendung durch leichtes Kippen oder Schwenken gemischt werden. Die Erzeugung von Schaum sollte dabei vermieden werden.
3. Wenn mit der Testdurchführung einmal begonnen ist, sollten alle nachfolgenden Schritte ohne Unterbrechung und innerhalb der empfohlenen Zeitgrenzen vollzogen werden.
4. Unmittelbar nach Entnahme der Reagenzien sollten die Flaschen wieder mit ihren Stopfen verschlossen werden. Die Verschlüsse dürfen nicht verwechselt werden!
5. Für jede Probe muss eine separate Einmal-Pipettenspitze verwendet werden, um eine Verschleppung bzw. Kreuzkontamination zu vermeiden.
6. Alle Proben und Standards sollten gleichzeitig getestet werden, so dass die Bedingungen für alle identisch sind.
7. Komponenten von verschiedenen Chargen dürfen nicht verwendet oder gemischt werden.
8. Reagenzien dürfen nach der Verfallszeit nicht benutzt werden.
9. Es sollten ständig die Präzision und die Richtigkeit für die Laborgeräte kontrolliert werden, die man für das Testverfahren benutzt (Mikropipetten, Washer, ELISA-Reader, etc.).

#### 4. GESUNDHEITS- UND SICHERHEITSVORSCHRIFTEN

1. Im Laboratorium darf nicht geraucht, gegessen und getrunken werden. Das Pipettieren mit dem Mund ist nicht zulässig.
2. Beim Einsatz von Proben menschlichen Ursprungs müssen Einmalhandschuhe getragen werden.
3. Der Kontakt der Stopp-Lösung mit Haut und Schleimhäuten sollte vermieden werden, da mögliche Reizungen, Verbrennungen oder Vergiftungsgefahr auftreten können. Sollte ein Kontakt entstehen, muss der betroffene Bereich mit ausreichend Wasser abgespült werden.
4. Die Handhabung und die Beseitigung von chemischen Produkten muss nach den Richtlinien für eine gute Laborpraxis (GLP) erfolgen.

#### 5. REAGENZIEN

Der Kit enthält Reagenzien für 96 Bestimmungen. Sie müssen bei 2-8°C gelagert werden. Das Verfallsdatum ist auf jeder Komponente sowie auf der Verpackung angegeben.

1. Mikrotiterplatte mit 96 Kavitäten (12 Streifen mit je 8 Vertiefungen, einzeln abbrechbar), beschichtet mit haselnussbindenden Antikörpern.
2. Haselnuss Standards: 5 Fläschchen mit je 1,0 mL (0, 1, 4, 10, 40 ppm Haselnuss), rot eingefärbt, gebrauchsfertig
3. Konjugat (anti-Haselnuss-Peroxidase), 15 mL, rot eingefärbt, gebrauchsfertig.
4. Substratlösung (TMB), 15 mL, gebrauchsfertig.
5. Stopp-Lösung (1 N saure Lösung), 15 mL, gebrauchsfertig.
6. Extraktions- und Verdünnungspuffer (TRIS), 2 x 120 mL als 10x-Konzentrat, rot eingefärbt. Gebrauchslösung: 1+9 mit dest. Wasser verdünnen. Der verdünnte Puffer ist bei 4°C mindestens eine Woche haltbar. Falls bei der gekühlten Lagerung Kristalle ausfallen sollten, das Konzentrat 15 Minuten im Wasserbad (37°C) erwärmen.
7. Waschlösung (PBS + Tween 20), 60 mL als 10x-Konzentrat. Gebrauchslösung: 1+9 mit dest. Wasser verdünnen. Der verdünnte Puffer ist bei 4°C mindestens 4 Wochen haltbar. Falls bei der gekühlten Lagerung Kristalle ausfallen sollten, das Konzentrat 15 Minuten im Wasserbad (37°C) erwärmen.
8. Ein wiederverschließbarer Plastikbeutel für die Lagerung unbenutzter Streifen.
9. Arbeitsanleitung.

#### 6. ZUSÄTZLICH BENÖTIGTE GERÄTE UND REAGENZIEN

##### Geräte

- 100 - 1000 µL Mikropipetten
- Messkolben
- Analysenwaage
- Mörser, Mixer
- Wasserbad
- Zentrifuge
- Mikrotiterplatten-Photometer (450 nm)

##### Reagenzien

- bidestilliertes Wasser



## 7. PROBENVORBEREITUNG

Aufgrund der hohen Gefahr von Kreuzkontaminationen müssen alle verwendeten Arbeitsgeräte wie Spatel, Mörser, Glasgefäße etc. vor und nach jeder Probe **gründlich gereinigt** werden. Haselnussproteine haften sehr stark an den Oberflächen und können in manchen Fällen selbst einer herkömmlichen Spülmaschinenreinigung standhalten. Es wird empfohlen, die Reihenfolge der Extraktionen festzuhalten, um eventuelle Kreuzkontaminationen durch Vorextrakte besser identifizieren zu können.

Folgende Probenvorbereitung sollte für alle Arten von Proben angewandt werden:

1. Um eine möglichst homogene und repräsentative Probennahme zu gewährleisten, sollten mindestens 5 g Probe in einem Mörser, Schlagmühle etc. fein zermahlen und gut durchgemischt werden.
2. Von dieser Mischung wird 1 g entnommen und in 20 mL **verdünntem** Extraktionspuffer suspendiert. Anschließend wird die Suspension für 15 min in einem vorgeheizten Wasserbad bei 60 °C inkubiert. Die Proben sollten währenddessen alle zwei Minuten geschüttelt werden, um eine gute Durchmischung zu gewährleisten.
3. Die Proben werden bei mindestens 2500 g für 10 min zentrifugiert. Sollte sich der Überstand anschließend nicht partikelfrei abtrennen lassen, kann gegebenenfalls noch einmal filtriert werden.
4. Für den Test werden pro Kavität 100 µL partikelfreie Lösung eingesetzt. Sollte das Ergebnis des Tests oberhalb des Messbereichs liegen, kann die Lösung gegebenenfalls mit **verdünntem** Extraktionspuffer weiter verdünnt und erneut bestimmt werden.

## 8. TESTDURCHFÜHRUNG

Der Waschpuffer liegt als 10faches Konzentrat vor und muss vor dem Gebrauch 1+9 mit bidestilliertem Wasser **verdünnt** werden.

Die **gebrauchsfertigen** Standards sollten in jedem Fall im Doppelansatz bestimmt werden. Bei größeren Mengen von Proben sollten die Standards einmal vor und einmal nach den Proben pipettiert und der Mittelwert zur Auswertung herangezogen werden.

Unter Berücksichtigung von GLP und Qualitätsmanagement ist eine Bestimmung der Proben im Doppelansatz zu empfehlen.

Die Testdurchführung verläuft nach dem folgenden Schema:

1. Proben nach Vorschrift vorbereiten.
2. 100 µL **gebrauchsfertige** Standards bzw. vorbehandelte Proben im Doppelansatz in die Vertiefungen der Mikrotiterplatte geben.
3. Platte 20 Minuten bei Raumtemperatur inkubieren.
4. Platte wie folgt dreimal waschen: Vertiefungen der Platte entleeren (auskippen oder absaugen) und 300 µL endverdünnte Waschlösung dazugeben. Nach der dritten Wiederholung Vertiefungen erneut leeren und Flüssigkeitsreste durch Ausschlagen der Mikrotiterplatte auf einem Papiertuch entfernen. Der Waschvorgang ist ein kritischer Schritt. Ungenügendes Waschen führt zu einer geringen Präzision und fälschlicherweise erhöhten Extinktionen.
5. 100 µL Konjugat (anti-Haselnuss-Peroxidase) in die Vertiefungen pipettieren.
6. Platte 20 Minuten bei Raumtemperatur inkubieren.
7. Vertiefungen wie in Punkt 4 beschrieben waschen.
8. 100 µL Substratlösung zugeben.
9. Platte abdecken und 20 Minuten bei Raumtemperatur im Dunkeln inkubieren.
10. Substratreaktion durch Zugabe von 100 µL Stopp-Lösung (1 N saure Lösung) beenden.
11. Nach sorgfältigem Mischen erfolgt die Messung der Extinktion bei 450 nm (eventuell Referenzwellenlänge 620 nm). Die Farbe ist maximal 30 Minuten stabil.

## 9. BERECHNUNG DER ERGEBNISSE

1. Für jede Probe bzw. jeden Standard wird die gemittelte Extinktion berechnet.
2. Aus den gemittelten Werten der Standardreihe wird graphisch auf halblogarithmischem Millimeterpapier oder über ein EDV-Programm (4-Parameter-Auswertung) eine Eichkurve erstellt. Die Extinktion jedes Standards (y-Achse) wird gegen die Konzentration in ppm (x-Achse) aufgetragen.
3. Mit Hilfe dieser Kurve wird für die gemittelten Extinktionen der Proben der Wert in ppm für Haselnuss abgelesen. Die ermittelten Konzentrationen beziehen sich direkt auf den Haselnussgehalt der eingewogenen Probe. Sollte der Extrakt aufgrund eines zu hohen Haselnussgehalts zusätzlich verdünnt worden sein, muss dies bei der Auswertung berücksichtigt werden.

**10. TYPISCHE STANDARDKURVE**

Die folgende Tabelle enthält ein Beispiel für eine Standardreihe. Die Werte dieser Standardkurve sind nur als Beispiel bestimmt und dürfen nicht an Stelle der in jedem Test neu zu erstellenden Kurve verwendet werden.

Haselnuss (ppm)	OD-% von 40 ppm
40	100
10	41
4	23
1	9
0	5

**11. TECHNISCHE DATEN**

**Empfindlichkeit**

Die mittlere untere Nachweisgrenze des **Haselnuss Tests** beträgt 0,3 ppm.

Die untere Bestimmungsgrenze des **Haselnuss Tests** beträgt 1 ppm.

Proben mit Ergebnissen kleiner der Bestimmungsgrenze sollten aufgrund der Vielfalt an Matrices und deren Einfluss auf den Hintergrund als negativ bewertet werden.

**Spezifität**

Für die folgenden Nahrungsmittel wurde ein negatives Ergebnis und damit keine Kreuzreaktivität festgestellt:

Weizen	Soja	Paranuss
Gerste	Mohn	Pistazie
Roggen	Sonnenblumenk.	Macadamianuss
Hafer	Kürbiskerne	Marone
Buchweizen	Pinienkerne	Kakao
Mais	Cashewnuss	Magermilchpulv.
Reis	Sesam	Gluten
Erbse	Erdnuss	Lecithin
Kichererbse	Walnuss	Gelatine
Weißer Bohne	Kokosnuss	Apfel
Mandel		

**Präzision**

Intra-Assay Präzision	7 – 12 %
Inter-Assay Präzision	3 – 13 %

**Linearität**

Die schrittweise Verdünnung dotierter Proben über fünf Stufen (Kekse, Cerealien, Eiscreme, Schokolade) ergab Verdünnungslinearitäten von 87 – 121%.

**Wiederfindung**









Die folgenden mittleren Wiederfindungsraten wurden anhand aufgestockter Proben bestimmt:

Kekse	101 %
Cerealien	99 %
Eiscreme	90 %
Schokolade	83 %

## 12. LITERATUR

1. Wigotzki M, et al. (2000) – Influence of varieties, storage and heat treatment on IgE-binding proteins in hazelnuts (*Corylus avellana*). *Food and Agric Immunol*, 12:217-29
2. Le TM, et al. (2008) – Reported food allergy to peanut, tree nuts and fruit: comparison of clinical manifestations, prescription of medication and impact on daily life. *Allergy*, 63(7):910-6
3. Luettkopf D, et al. (2001) – Comparison of four variants of major allergen in hazelnut (*Corylus avellana*) Cor a 1.04 with the major hazel pollen allergen Cor a 1.01. *Mol Immunol*, 38:515-25
4. Ballmer-Weber BK, et al. (2008) – Predictive value of the sulfidoleukotriene release assay in oral allergy syndrome to celery, hazelnut, and carrot. *J Investig Allergol Clin Immunol*, 18(2):93-99
5. Ben Rejeb S, et al. (2005) – Multi-allergen screening immunoassay for the detection of protein markers of peanut and four tree nuts in chocolate. *Food Addit Contam*, 22(8):709-15
6. Lauer I, et al. (2004) – Hazelnut (*Corylus avellana*) vicilin Cor a 11: molecular characterization of a glycoprotein and its allergenic activity. *Biochem J*, 383:327-334
7. Holzhauser T, et al. (2002) – Detection of potentially allergenic hazelnut (*Corylus avellana*) residues in food: a comparative study with DNA PCR-ELISA and protein sandwich-ELISA. *J Agric Food Chem*, 50(21):5808-15
8. Kiening M, et al. (2005) – Sandwich immunoassays for the determination of peanut and hazelnut traces in foods. *J Agric Food Chem*, 53(9):3321-7

## SYMBOLS USED WITH DEMEDITEC ASSAYS

Symbol	English	Deutsch	Français	Español	Italiano
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las instrucciones de uso	Consultare le istruzioni per l'uso
	European Conformity	CE-Konformitätskennzeichnung	Conformité aux normes européennes	Conformidad europea	Conformità europea
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Numéro de catalogue	Número de catálogo	Numero di Catalogo
	Lot. No. / Batch code	Chargen-Nr.	Numéro de lot	Número de lote	Numero di lotto
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeitsdatum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
Distributed by	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore
Content	Content	Inhalt	Conditionnement	Contenido	Contenuto
Volume/No.	Volume / No.	Volumen/Anzahl	Volume/Quantité	Volumen/Número	Volume/Quantità