

Product information

Information about other products is available at: www.demeditec.com



User's Manual

Dengue Virus IgG ELISA

Enzyme immunoassay for the qualitative determination of IgG-class antibodies against Dengue Virus in human serum or plasma

IVD

CE

REF

DENG0120



96

ENGLISH**1. INTRODUCTION**

Dengue virus is a single-stranded RNA virus of about 50 nm in diameter belonging to the genus Flavivirus. Dengue and dengue hemorrhagic fever are caused by one of four closely related, but antigenically distinct, virus serotypes (DEN-1, DEN-2, DEN-3, and DEN-4). Infection with one of these serotypes does not provide crossprotective immunity, so persons living in a dengue-endemic area can have four dengue infections during their lifetimes. The viruses are transmitted by Aedes aegypti, a domestic, day-biting mosquito that prefers to feed on humans. Infection with dengue viruses produces a spectrum of clinical illness ranging from a nonspecific viral syndrome to severe and fatal hemorrhagic disease. It is primarily a disease of the tropics; its global distribution is comparable to that of malaria, and an estimated 2.5 billion people live in areas at risk for epidemic transmission. - Globally, there are an estimated 50 to 100 million cases of dengue fever and several hundred thousand cases of dengue hemorrhagic fever.

- The case-fatality rate of DHF in most countries is about 5%; most fatal cases are among children and young adults.
- Important risk factors for DHF include the strain and serotype of the infecting virus, as well as the age, immune status, and
- genetic predisposition of the patient.
- Risk groups: residents of or visitors to tropical urban areas.

Species	Disease	Symptoms	Mechanism of Infection
Dengue virus	Dengue Dengue hemorrhagic fever (DHF) or Breakbone fever	Sudden onset of fever, severe headache, myalgias and arthralgia leukopenia, thrombocytopenia and hemorrhagic manifestations	Transmission by mosquitos (Aedes aegypti)

The presence of virus resp. infection may be identified by

- Serology: Detection of antibodies by ELISA

Infection produces lifelong immunity, but the antigenically distinct serotypes do not provide cross-protective immunity, so a person can theoretically experience four dengue infections; a dengue vaccine is not available

2. INTENDED USE

The Demeditec Dengue virus IgG-ELISA is intended for the qualitative determination of IgG class antibodies against Dengue virus in human serum or plasma (citrate).

3. PRINCIPLE OF THE ASSAY

The qualitative immunoenzymatic determination of IgG-class antibodies against Dengue Virus is based on the ELISA (Enzyme-linked Immunosorbent Assay) technique.

Microtiter strip wells are precoated with Dengue Virus antigens type 2 to bind corresponding antibodies of the specimen. After washing the wells to remove all unbound sample material horseradish peroxidase (HRP) labelled anti-human IgG conjugate is added. This conjugate binds to the captured Dengue Virus-specific antibodies. The immune complex formed by the bound conjugate is visualized by adding Tetramethylbenzidine (TMB) substrate which gives a blue reaction product. The intensity of this product is proportional to the amount of Dengue Virus-specific IgG antibodies in the specimen. Sulphuric acid is added to stop the reaction. This produces a yellow endpoint colour. Absorbance at 450 nm is read using an ELISA microwell plate reader.

4. MATERIALS

4.1. Reagents supplied

- **Dengue Virus Coated Wells (IgG):** 12 breakapart 8-well snap-off strips coated with Dengue Virus Type 2 antigen; in resealable aluminium foil.
- **IgG Sample Diluent ***:** 1 bottle containing 100 ml of buffer for sample dilution; pH 7.2 ± 0.2; coloured yellow; ready to use; white cap.
- **Stop Solution:** 1 bottle containing 15 ml sulphuric acid, 0.2 mol/l; ready to use; red cap.
- **Washing Solution (20x conc.)*:** 1 bottle containing 50 ml of a 20-fold concentrated buffer (pH 7.2 ± 0.2) for washing the wells; white cap.
- **Dengue Virus anti-IgG Conjugate**:** 1 bottle containing 20 ml of peroxidase labelled rabbit antibody to human IgG; coloured blue, ready to use; black cap.
- **TMB Substrate Solution:** 1 bottle containing 15 ml 3,3',5,5'-tetramethylbenzidine (TMB); ready to use; yellow cap.
- **Dengue Virus IgG Positive Control***:** 1 bottle containing 2 ml; coloured yellow; ready to use; red cap.
- **Dengue Virus IgG Cut-off Control***:** 1 bottle containing 3 ml; coloured yellow; ready to use; green cap.
- **Dengue Virus IgG Negative Control***:** 1 bottle containing 2 ml; coloured yellow; ready to use; blue cap.

* contains 0.1 % Bronidox L after dilution

** contains 0.2 % Bronidox L

*** contains 0.1 % Kathon

4.2. Materials supplied

- 1 Strip holder
- 1 Cover foil
- 1 Test protocol
- 1 distribution and identification plan

4.3. Materials and Equipment needed

- ELISA microwell plate reader, equipped for the measurement of absorbance at 450/620 nm
- Incubator 37°C
- Manual or automatic equipment for rinsing wells
- Pipettes to deliver volumes between 10 and 1000 µl
- Vortex tube mixer
- Deionised or (freshly) distilled water
- Disposable tubes
- Timer

5. STABILITY AND STORAGE

The reagents are stable up to the expiry date stated on the label when stored at 2...8 °C.

6. REAGENT PREPARATION

It is very important to bring all reagents, samples and controls to room temperature (20...25 °C) before starting the test run!

6.1. Coated snap-off Strips

The ready to use breakapart snap-off strips are coated with Dengue Virus Type 2 antigen. Store at 2...8 °C. *Immediately after removal of strips, the remaining strips should be resealed in the aluminium foil along with the desiccant supplied and stored at 2...8 °C; stability until expiry date.*

6.2. Dengue Virus anti-IgG Conjugate

The bottle contains 20 ml of a solution with anti-human-IgG horseradish peroxidase, buffer, stabilizers, preservatives and an inert blue dye. The solution is ready to use. Store at 2...8 °C. *After first opening stability until expiry date when stored at 2...8 °C.*

6.3. Controls

The bottles labelled with Positive, Cut-off and Negative Control contain a ready to use control solution. It contains 0.1% Kathon and has to be stored at 2...8 °C. *After first opening stability until expiry date when stored at 2...8 °C.*

6.4. IgG Sample Diluent

The bottle contains 100 ml phosphate buffer, stabilizers, preservatives and an inert yellow dye. It is used for the dilution of the patient specimen. This ready to use solution has to be stored at 2...8 °C. *After first opening stability until expiry date when stored at 2...8 °C.*

6.5. Washing Solution (20xconc.)

The bottle contains 50 ml of a concentrated buffer, detergents and preservatives. Dilute washing solution 1+19; e.g. 10 ml washing solution + 190 ml fresh and germ free redistilled water. The diluted buffer is stable for 5 days at room temperature. *Crystals in the solution disappear by warming up to 37 °C in a water bath. After first opening the concentrate is stable until the expiry date.*

6.6. TMB Substrate Solution

The bottle contains 15 ml of a tetramethylbenzidine/hydrogen peroxide system. The reagent is ready to use and has to be stored at 2...8 °C, away from the light. *The solution should be colourless or could have a slight blue tinge. If the substrate turns into blue, it may have become contaminated and should be thrown away. After first opening stability until expiry date when stored at 2...8 °C.*

6.7. Stop Solution

The bottle contains 15 ml 0.2 M sulphuric acid solution (R 36/38, S 26). This ready to use solution has to be stored at 2...8 °C.

After first opening stability until expiry date.

7. SPECIMEN COLLECTION AND PREPARATION

Use human serum or plasma (citrate) samples with this assay. If the assay is performed within 5 days after sample collection, the specimen should be kept at 2...8 °C; otherwise they should be aliquoted and stored deep-frozen (-20 to -70 °C). If samples are stored frozen, mix thawed samples well before testing. *Avoid repeated freezing and thawing.*

Heat inactivation of samples is not recommended.

7.1. Sample Dilution

Before assaying, all samples should be diluted 1+100 with IgG Sample Diluent. Dispense 10 µl sample and 1 ml IgG Sample Diluent into tubes to obtain a 1+100 dilution and thoroughly mix with a Vortex.

8. ASSAY PROCEDURE

8.1. Test Preparation

Please read the test protocol carefully **before** performing the assay. Result reliability depends on strict adherence to the test protocol as described. The following test procedure is only validated for manual procedure. If performing the test on ELISA automatic systems we recommend to increase the washing steps from three to five and the volume of washing solution from 300µl to 350µl to avoid washing effects. Prior to commencing the assay, the distribution and identification plan for all specimens and controls should be carefully established on the result sheet supplied in the kit. Select the required number of microtiter strips or wells and insert them into the holder.

Please allocate at least:

- | | |
|----------------------|-----------------------------|
| 1 well (e.g. A1) | for the substrate blank, |
| 1 well (e.g. B1) | for the negative control, |
| 2 wells (e.g. C1+D1) | for the cut-off control and |
| 1 well (e.g. E1) | for the positive control. |

It is recommended to determine controls and patient samples in duplicate, if necessary.

Perform all assay steps in the order given and without any appreciable delays between the steps.

A clean, disposable tip should be used for dispensing each control and sample.

Adjust the incubator to 37° ± 1°C.

1. Dispense 100µl controls and diluted samples into their respective wells. Leave well A1 for substrate blank.
2. Cover wells with the foil supplied in the kit.
3. **Incubate for 1 hour ± 5 min at 37±1 °C.**
4. When incubation has been completed, remove the foil, aspirate the content of the wells and wash each well three times with 300µl of Washing Solution. Avoid overflows from the reaction wells. The soak time between each wash cycle should be >5 sec. At the end carefully remove remaining fluid by tapping strips on tissue paper prior to the next step!

Note: Washing is critical! Insufficient washing results in poor precision and falsely elevated absorbance values.

5. Dispense 100µl Dengue Virus anti-IgG Conjugate into all wells except for the blank well (e.g. A1). Cover with foil.
6. **Incubate for 30 min at room temperature. Do not expose to direct sunlight.**
7. Repeat step 4.
8. Dispense 100µl TMB Substrate Solution into all wells
9. **Incubate for exactly 15 min at room temperature in the dark.**

10. Dispense 100µl Stop Solution into all wells in the same order and at the same rate as for the TMB Substrate Solution.

Any blue colour developed during the incubation turns into yellow.

Note: Highly positive patient samples can cause dark precipitates of the chromogen! These precipitates have an influence when reading the optical density. Predilution of the sample with physiological sodium chloride solution, for example 1+1, is recommended. Then dilute the sample 1+100 with dilution buffer and multiply the results in U by 2.

11. Measure the absorbance of the specimen at 450/620 nm within 30 min after addition of the Stop Solution.

8.2. Measurement

Adjust the ELISA Microwell Plate Reader **to zero** using the **substrate blank in well A1**.

If - due to technical reasons - the ELISA reader cannot be adjusted to zero using the substrate blank in well A1, subtract the absorbance value of well A1 from all other absorbance values measured in order to obtain reliable results!

Measure the absorbance of all wells at **450 nm** and record the absorbance values for each control and patient sample in the distribution and identification plan.

Dual wavelength reading using 620 nm as reference wavelength is recommended.

Where applicable calculate the **mean absorbance values** of all duplicates.

9. RESULTS

9.1. Run Validation Criteria

In order for an assay to be considered valid, the following criteria must be met:

- **Substrate blank** in A1: Absorbance value < 0.100.
- **Negative control** in B1: Absorbance value < 0.200 and < cut-off
- **Cut-off control** in C1 and D1: Absorbance value 0.150 – 1.30.
- **Positive control** in E1: Absorbance value > cut-off.

If these criteria are not met, the test is not valid and must be repeated.

9.2. Calculation of Results

The cut-off is the mean absorbance value of the Cut-off control determinations.

*Example: Absorbance value Cut-off control 0.49 + absorbance value Cut-off control 0.47 = 0.96 / 2 = 0.48
Cut-off = 0.48*

9.3. Interpretation of Results

Samples are considered **POSITIVE** if the absorbance value is higher than 10% over the cut-off.

Samples with an absorbance value of 10% above or below the cut-off should not be considered as clearly positive or negative
→ **grey zone**

It is recommended to repeat the test again 2 - 4 weeks later with a fresh sample. If results in the second test are again in the grey zone the sample has to be considered **NEGATIVE**.

Samples are considered **NEGATIVE** if the absorbance value is lower than 10% below the cut-off.

9.3.1. Results in Demeditec Units

$$\frac{\text{Patient (mean) absorbance value} \times 10}{\text{Cut-off}} = [\text{Demeditec-Units} = U]$$

Example: $\frac{1.786 \times 10}{0.48} = 37 \text{ U (Demeditec Units)}$

Cut-off:	10	U
Grey zone:	9-11	U
Negative:	<9	U
Positive:	>11	U

10. SPECIFIC PERFORMANCE CHARACTERISTICS

10.1. Precision

Interassay	n	Mean	Cv (%)
Pos. Serum	8	0.47	6.76
Intraassay	n	Mean	Cv (%)
Pos Serum	8	0.98	4.34

10.2. Diagnostic Specificity

The diagnostic specificity is defined as the probability of the assay of scoring negative in the absence of the specific analyte.

It is 93 %.

10.3. Diagnostic Sensitivity

The diagnostic sensitivity is defined as the probability of the assay of scoring positive in the presence of the specific analyte.

It is >90 %.

10.4. Interferences

Interferences with hemolytic, lipemic or icteric sera are not observed up to a concentration of 10 mg/ml hemoglobin, 5 mg/ml triglycerides and 0.2 mg/ml bilirubin.

Note: The results refer to the groups of samples investigated; these are not guaranteed specifications.

11. LIMITATIONS OF THE PROCEDURE

The **Flaviviridae family** includes the serotypes of Dengue virus as well as the **yellow fever, Japanese encephalitis** viruses and **Tick borne encephalitis (TBE)**. There is a cross reactivity among flaviviruses, due to the presence of common antigen determinants. Diagnosis of Dengue infection should be made in conjunction with other clinical signs and symptoms and laboratory findings. Known cross reactions among Dengue antigens must be considered during interpretation since epitopes are known to react with other flaviviruses. It is recommended to test a reactive sample with a second serological method. Bacterial contamination or repeated freeze-thaw cycles of the specimen may affect the absorbance values. Diagnosis of an infectious disease should not be established on the basis of a single test result. A precise diagnosis should take into consideration clinical history, symptomatology as well as serological data. In immuno-compromised patients and newborns serological data only have restricted value.

12. PRECAUTIONS AND WARNINGS

- In compliance with article 1 paragraph 2b European directive 98/79/EC the use of the in vitro diagnostic medical devices is intended by the manufacturer to secure suitability, performances and safety of the product. Therefore the test procedure, the information, the precautions and warnings in the instructions for use have to be strictly followed. The use of the testkits with analyzers and similar equipment has to be validated. Any change in design, composition and test procedure as well as for any use in combination with other products not approved by the manufacturer is not authorized; the user himself is responsible for such changes. The manufacturer is not liable for false results and incidents for these reasons. The manufacturer is not liable for any results by visual analysis of the patient samples.
- Only for in-vitro diagnostic use.
- All components of human origin used for the production of these reagents have been tested for anti-HIV antibodies, anti-HCV antibodies and HBsAg and have been found to be non-reactive. Nevertheless, all materials should still be regarded and handled as potentially infectious.
- Do not interchange reagents or strips of different production lots.
- No reagents of other manufacturers should be used along with reagents of this test kit.
- Do not use reagents after expiry date stated on the label.
- Use only clean pipette tips, dispensers, and lab ware.
- Do not interchange screw caps of reagent vials to avoid cross-contamination.
- Close reagent vials tightly immediately after use to avoid evaporation and microbial contamination.
- After first opening and subsequent storage check conjugate and control vials for microbial contamination prior to further use.
- To avoid cross-contamination and falsely elevated results pipette patient samples and dispense conjugate without splashing accurately to the bottom of wells.
- The ELISA is only designed for qualified personnel who are familiar with good laboratory practice.

WARNING:	In the used concentration Bronidox L has hardly any toxicological risk upon contact with skin and mucous membranes!
WARNING:	Sulphuric acid irritates eyes and skin. Keep out of the reach of children. Upon contact with the eyes, rinse thoroughly with water and consult a doctor!

12.1. Disposal Considerations

Residues of chemicals and preparations are generally considered as hazardous waste. The disposal of this kind of waste is regulated through national and regional laws and regulations. Contact your local authorities or waste management companies which will give advice on how to dispose hazardous waste.

DEUTSCH**1. EINLEITUNG**

Dengue-Viren sind umhüllte Einzelstrang RNA Viren von etwa 50 nm Durchmesser. Sie gehören zur Familie der Flaviviridae. Man unterscheidet vier Serotypen (DEN-1, DEN-2, DEN-3 und DEN-4). Infektion mit einem Serotyp erzeugt keine Immunität gegenüber den anderen Serotypen.

Dengue-Viren sind weit verbreitet. In Süd- und Mittelamerika, Westafrika, Südostasien und im West pazifischen Ozean Dengue-Fieber endemisch. Reservoir der Viren ist der Mensch. Die Übertragung erfolgt durch Aedes aegypti. Klinisch können drei Krankheitsbilder unterschieden werden:

- das Dengue-Fieber, das sich nach einer Inkubationszeit von 1-2 Wochen mit Schüttelfrost (bis 40 °C Fieber), Kopf-, Glieder- und Muskelschmerzen manifestiert. Die Krankheit ist insgesamt gutartig.
- das hämorrhagische Dengue-Fieber, das durch Haut- und Organblutungen imponiert, verläuft weitaus schwerer. Klinisch finden sich Petechien, starkes Nasenbluten, Bluterbrechen, Meläna und Hämaturien. Eine weitere Steigerung des Krankheitsbildes führt zum
- Dengue-Schocksyndrom. Es kommt zu massiven Organblutungen, die auch das ZNS erfassen (Massenblutung ins Gehirn) und dann häufig (10-40 %) letal enden.

Wegen der unterschiedlich schweren Verlaufsformen variiert die Letalität erheblich. Im Mittel liegt sie bei 1-3 %, kann bei den schweren Formen jedoch bis auf 80 % ansteigen. Eine kausale Therapie ist nicht möglich, eine spezielle Prophylaxe existiert nicht.

Spezies	Erkrankung	Symptome	Infektionsmodus
Dengue-Virus	Dengue-Fieber hämorrhagisches Dengue-Fieber Dengue- Schocksyndrom Cut	Schüttelfrost, Fieber, Kopf-, Glieder- und Muskelschmerzen Petechien, Nasenbluten, Bluterbrechen, Meläna, Hämaturien, Organblutungen	Übertragung durch Mückenstich (Aedes aegypti)

Infektionen können bestätigt werden durch:

- Serologie: Nachweis spezifischer Antikörper mittels ELISA

2. VERWENDUNGSZWECK

Der Demeditec Dengue-Virus IgG ELISA ist für den qualitativen Nachweis spezifischer IgG-Antikörper gegen Dengue-Viren (1-4) in humanem Serum oder Plasma (Citrat) bestimmt.

3. TESTPRINZIP

Die qualitative immunenzymatische Bestimmung von spezifischem IgG gegen Dengue-Viren beruht auf der ELISA (Enzyme-linked Immunosorbent Assay)-Technik. Mikrotiterstreifen als solide Phase sind beschichtet mit Dengue-Virus Antigen. (Type 2) Vorhandene spezifische Antikörper in der Probe binden an die immobilisierten Antigene der Mikrotiterplatte. Meerrettich-Peroxidase (HRP) -konjugierte antihuman IgG-Antikörper binden an Antigen-Antikörperkomplexe in positiven Proben. Die entstandenen Immunkomplexe werden durch Blaufärbung nach Inkubation mit Tetramethylbenzidine (TMB) -Substratlösung nachgewiesen. Stoppen der enzymatischen Reaktion mit Schwefelsäure führt zu einem Farbumschlag von blau zu gelb, der einfach nachgewiesen und mit einem ELISA-Reader bei 450 nm gemessen werden kann.

4. MATERIALIEN

4.1. Mitgelieferte Reagenzien

- **Dengue-Virus beschichtete Mikrotiterstreifen (IgG):** 12 teilbare 8er-Streifen, beschichtet mit Dengue-Virus Antigen; in wiederverschließbarem Aluminiumbeutel.
- **IgG-Probenverdünnungspuffer***:** 1 Flasche mit 100 ml Puffer zur Probenverdünnung; pH 7.2 ± 0.2; gelb gefärbt; gebrauchsfertig; weiße Verschlusskappe.
- **Stopplösung:** 1 Fläschchen mit 15 ml Schwefelsäure, 0.2 mol/l, gebrauchsfertig; rote Verschlusskappe.
- **Waschlösung (20x konz.)*:** 1 Flasche mit 50 ml eines 20-fach konzentrierten Puffers zum Waschen der Kavitäten; pH 7.2 ± 0.2; weiße Verschlusskappe.
- **Dengue-Virus anti-IgG-Konjugat**:** 1 Flasche mit 20 ml Peroxidase-konjugierten Antikörpern gegen humanes IgG; blau gefärbt; gebrauchsfertig; schwarze Verschlusskappe.
- **TMB-Substratlösung:** 1 Fläschchen mit 15 ml 3,3',5,5'-Tetramethylbenzidin (TMB); gebrauchsfertig; gelbe Verschlusskappe.
- **Dengue-Virus IgG Positivkontrolle***:** 1 Fläschchen mit 2 ml; gelb gefärbt; rote Verschlusskappe; gebrauchsfertig.
- **Dengue-Virus IgG Cut-off Kontrolle***:** 1 Fläschchen mit 3 ml; gelb gefärbt; grüne Verschlusskappe; gebrauchsfertig.
- **Dengue-Virus IgG Negativkontrolle***:** 1 Fläschchen mit 2 ml; gelb gefärbt; blaue Verschlusskappe; gebrauchsfertig.

* enthält 0.1 % Bronidox L nach Verdünnung

** enthält 0.2 % Bronidox L

*** enthält 0.1 % Kathon

4.2. Mitgeliefertes Zubehör

- 1 selbstklebende Abdeckfolie
- 1 Rahmenhalter
- 1 Arbeitsanleitung
- 1 Ergebnisblatt

4.3. Erforderliche Materialien und Geräte

- Photometer mit Filtern 450/620 nm
- Feuchtkammer/Brutschrank mit Thermostat
- Manuelle oder automatische Waschvorrichtung
- Mikropipetten mit Einmalspitzen (10, 100, 200, 1000 µl)
- Vortex-Mischer
- Plastikröhrchen für den einmaligen Gebrauch
- Röhrchen-Ständer
- Aqua dest.
- Timer

5. STABILITÄT UND LAGERUNG

Testkit bei 2...8°C lagern. Die Reagenzien nicht nach den angegebenen Verfallsdaten verwenden. Die Verfallsdaten sind jeweils auf den Flaschenetiketten und auf dem Außenetikett angegeben.

6. VORBEREITUNG DER REAGENZIEN

Alle Reagenzien, Proben und Kontrollen sind vor ihrer Verwendung auf Raumtemperatur (20...25 °C) zu bringen!

6.1. Beschichtete Streifen

Die abbrechbaren Streifen sind mit inaktiviertem Dengue-Virus Antigen beschichtet. Die gebrauchsfertigen Vertiefungen sind bei 2...8 °C aufzubewahren. *Nichtverbrauchte Vertiefungen im Aluminiumbeutel zusammen mit dem Trockenmittel sofort wieder verschließen und bei 2...8 °C lagern. Haltbarkeit bis zum angegebenen Verfallsdatum.*

6.2. Dengue-Virus anti-IgG-Konjugat

Die Flasche enthält 20 ml einer Lösung von anti-human IgG-Meerrettichperoxidase, Puffer, Stabilisatoren, Konservierungsmittel und einen inerten blauen Farbstoff. Die gebrauchsfertige Lösung ist bei 2...8 °C aufzubewahren. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2...8 °C).*

6.3. Kontrollen

Die Fläschchen mit Kontrollen enthalten 2 bzw. 3 ml gebrauchsfertige Kontrolllösung. Die gebrauchsfertigen Lösungen sind bei 2...8 °C aufzubewahren und enthalten 0.1 % Kathon. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2...8 °C).*

6.4. IgG-Probenverdünnungspuffer

Die Flasche enthält 100 ml Phosphatpuffer, Stabilisatoren, Konservierungsmittel und einen inerten gelben Farbstoff. Die gebrauchsfertige Lösung ist bei 2...8 °C aufzubewahren. Die Lösung wird für die Verdünnung der Proben eingesetzt. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2...8 °C).*

6.5. Waschlösung (20x konz.)

Die Flasche enthält 50 ml konzentrierten Puffer, Detergenzien und Konservierungsmittel. Der Inhalt wird auf einen Liter mit Aqua dest. verdünnt (1+19). Der verdünnte Puffer ist bei Raumtemperatur 5 Tage haltbar. Die Waschlösung wird zum Waschen der Streifen eingesetzt. *Sollte eine Kristallisation im Konzentrat auftreten, die Waschlösung auf 37 °C erwärmen und vor dem Verdünnen gut mischen. Nach dem ersten Öffnen, Konzentrat haltbar bis zum angegebenen Verfallsdatum (bei 2...8 °C).*

6.6 TMB-Substrat

Das Fläschchen enthält 15 ml eines Tetramethylbenzidin/Wasserstoffperoxidgemisches. Die gebrauchsfertige Lösung ist bei 2...8 °C vor Licht geschützt aufzubewahren. *Die Lösung ist leicht bläulich. Sollte die TMB-Substratlösung dunkelblau sein, ist sie kontaminiert und kann nicht im Test verwendet werden. Nach dem ersten Öffnen haltbar bis zum Verfallsdatum bei sachgerechter Lagerung von 2...8 °C.*

6.7. Stopplösung

Das Fläschchen enthält 15 ml 0,2 M Schwefelsäure (R36/38, S26). Die gebrauchsfertige Lösung ist bei 2...8 °C aufzubewahren. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2...8 °C).*

7. ENTNAHME UND VORBEREITUNG DER PROBEN

Es sollten humane Serum- oder Plasmaproben (Citrat) verwendet werden. Werden die Bestimmungen innerhalb von 5 Tagen nach Blutentnahme durchgeführt, können die Proben bei 2...8 °C aufbewahrt werden, sonst tiefgefrieren (-70...-20 °C). Wiederaufgetaute Proben vor dem Verdünnen gut schütteln. *Wiederholtes Tiefgefrieren und Auftauen vermeiden!*
Hitzeinaktivierung der Proben wird nicht empfohlen.

7.1. Probenverdünnung

Proben vor Testbeginn im Verhältnis 1+100 mit IgG-Probenverdünnungspuffer verdünnen, z.B. 10µl Probe und 1 ml IgG-Probenverdünnungspuffer in die entsprechenden Röhrchen pipettieren, um eine Verdünnung von 1+100 zu erhalten; gut mischen (Vortex).

8. TESTDURCHFÜHRUNG

8.1 Testvorbereitung

Gebrauchsinformation **vor** Durchführung des Tests sorgfältig lesen. Für die Zuverlässigkeit der Ergebnisse ist es notwendig, die Arbeitsanleitung genau zu befolgen. Die folgende Testdurchführung ist für die manuelle Methode validiert. Beim Arbeiten mit ELISA Automaten empfehlen wir, um Wascheffekte auszuschließen, die Zahl der Waschschrifte von drei auf fünf und das Volumen der Waschlösung von 300 µl auf 350 µl zu erhöhen. Vor Testbeginn auf dem mitgelieferten Ergebnisblatt die Verteilung bzw. Position der Patientenproben und Standards auf den Mikrotiterstreifen genau festlegen. Die benötigte Anzahl von Mikrotiterstreifen (Kavitäten) in den Streifenhalter einsetzen.

Hierbei mindestens

- | | |
|-----------------------------|-------------------------------------|
| 1 Vertiefung (z.B. A1) | für den Substratleerwert (Blank), |
| 1 Vertiefung (z.B. B1) | für die Negativ Kontrolle und |
| 2 Vertiefungen (z.B. C1+D1) | für die Cut-off Kontrolle und |
| 1 Vertiefung (z.B. E1) | für die Positiv Kontrolle vorsehen. |

Prinzipien der Qualitätssicherung in der Laboratoriumsmedizin erfordern zur höheren Sicherheit für Kontrollen und Patientenproben mindestens Doppelbestimmungen.

Den Test in der angegebenen Reihenfolge und ohne Verzögerung durchführen.

Für jeden Pipettierschritt der Kontrollen und Proben saubere Einmalspitzen verwenden.

Den Brutschrank auf 37 ± 1 °C einstellen.

1. Je 100 µl Kontrollen und vorverdünnte Proben in die entsprechenden Vertiefungen pipettieren. Vertiefung A1 ist für den Substratleerwert vorgesehen.
2. Die Streifen mit der mitgelieferten Abdeckfolie bedecken.
3. **1 h ± 5 min bei 37 °C inkubieren.**
4. Am Ende der Inkubationszeit Abdeckfolie entfernen und die Inkubationsflüssigkeit aus den Teststreifen absaugen. Anschließend dreimal mit 300µl Waschlösung waschen. Überfließen von Flüssigkeit aus den Vertiefungen vermeiden. Intervall zwischen Waschen und Absaugen sollte mindestens 5 sec betragen. Nach dem Waschen die Teststreifen mit den Öffnungen nach unten kurz auf Fliesspapier ausklopfen, um die restliche Flüssigkeit zu entfernen.

Beachte: Der Waschvorgang ist wichtig, da unzureichendes Waschen zu schlechter Präzision und falsch erhöhten Messergebnissen führt!

5. 100µl Dengue-Virus anti-IgG-Konjugat in alle Vertiefungen, mit Ausnahme der für die Berechnung des Leerwertes vorgesehenen, pipettieren. Mit Folie abdecken.
6. **30 min bei Raumtemperatur (20...25 °C) inkubieren. Nicht dem direkten Sonnenlicht aussetzen.**
7. Waschvorgang gemäß Punkt 4 wiederholen.
8. 100µl TMB-Substratlösung in alle Vertiefungen pipettieren.
9. **Genau 15 min im Dunkeln bei Raumtemperatur (20...25 °C) inkubieren.**
10. In alle Vertiefungen 100µl Stopplösung in der gleichen Reihenfolge und mit den gleichen Zeitintervallen wie bei der TMB-Substratlösung Zugabe pipettieren. *Während der Inkubation gebildete blaue Farbe schlägt in gelb um.*

Hinweis: Hochpositive Patientenproben können schwärzliche Präzipitate des Chromogens verursachen! Diese Präzipitate beeinflussen die Messwerte. Es wird empfohlen, die Patientenprobe mit physiologischer Kochsalzlösung 1 + 1 zu verdünnen und anschließend die verdünnte Probe mit IgG-Probenverdünnungspuffer 1 + 100 für den Test vorzubereiten. Das Ergebnis in U wird in diesem Fall mit zwei multipliziert.

11. Die Extinktion der Lösung in jeder Vertiefung bei 450/620 nm innerhalb von 30 min nach Zugabe der Stopplösung messen

8.2. Messung

Mit Hilfe des Substratleerwertes (Blank) in A1 den **Nullabgleich** des Mikrotiterplatten-Photometers (ELISA-Readers) vornehmen.

Falls diese Eichung aus technischen Gründen nicht möglich ist, muss nach der Messung der Extinktionswert der Position A1 von allen anderen Extinktionswerten abgezogen werden, um einwandfreie Ergebnisse zu erzielen!

Extinktion aller Kavitäten bei **450 nm** messen und die Messwerte der Kontrollen und Proben in das Ergebnisblatt eintragen.

Eine **bichromatische** Messung mit der Referenzwellenlänge **620 nm** wird empfohlen.

Falls Doppel- oder Mehrfachbestimmungen durchgeführt wurden, den **Mittelwert der Extinktionswerte** berechnen.

9. BERECHNUNG DER ERGEBNISSE

9.1. Testgültigkeitskriterien

Der Test wurde richtig durchgeführt, wenn er folgende Kriterien erfüllt:

- **Substrat-Leerwert** in A1: Extinktion < **0,100**
- **Negativ Kontrolle** in B1: Extinktion < **0,200 und < cut-off**
- **Cut-off Kontrolle** in C1 und D1: Extinktionswerte **0,150 – 1,300**
- **Positiv Kontrolle** in E1: Extinktionswerte > **Cut-off**

Sind diese Kriterien nicht erfüllt, ist der Testlauf ungültig und muss wiederholt werden.

9.2. Messwertberechnung

Der Cut-off ergibt sich aus dem Mittelwert der gemessenen Extinktionen der beiden Cut-off Kontrollen.

Beispiel: $0.47 \text{ OD Cut-off Kontrolle} + 0.49 \text{ OD Cut-off Kontrolle} = 0.96 : 2 = \underline{0.48}$
Cut-off = 0.48

9.3. Interpretation der Ergebnisse

Patientenproben gelten als **positiv**, wenn der Extinktionswert mindestens 10 % höher liegt als der Cut-Off.

Patientenproben mit Extinktionswerten 10 % über bzw. unter dem Cut-Off können nicht eindeutig als positiv bzw. negativ angesehen werden → **Grauzone**

Es wird empfohlen den Test nach 2 bis 4 Wochen mit einer frischen Patientenprobe zu wiederholen. Finden sich die Ergebnisse erneut innerhalb der Grauzone, gilt die Probe als **negativ**.

Patientenproben gelten als **negativ**, wenn der Extinktionswert mindestens 10 % unterhalb des Cut-Offs liegt.

Hinweis: Da Kreuzreaktionen mit anderen Flaviviren nicht ausgeschlossen werden können wird empfohlen, positive Testergebnisse mit einem zweiten Test zu bestätigen.

9.3.1. Ergebnisse in Demeditec-Einheiten [U]

Mittlere Extinktion der Patientenprobe x 10 = [Demeditec-Einheiten = U]
Cut-Off

Beispiel: $\frac{1.786 \times 10}{0.48} = 37 \text{ U (Demeditec Units)}$

Cut-Off:	10	U
Grauzone:	9-11	U
Negativ:	<9	U
Positiv:	>11	U

10. TESTMERKMALE

10.1. Präzision

Interassay	n	Mittelwert	Vk (%)
Pos. Serum	8	0.47	6.76

Intraassay	n	Mittelwert	Vk (%)
Pos Serum	8	0.98	4.34

10.2. Diagnostische Spezifität

Die diagnostische Spezifität ist definiert als die Wahrscheinlichkeit des Tests, ein negatives Ergebnis bei Fehlen des spezifischen Analyten zu liefern. Sie beträgt 93 %.

10.3. Diagnostische Sensitivität

Die diagnostische Sensitivität ist definiert als die Wahrscheinlichkeit des Tests, ein positives Ergebnis bei Vorhandensein des spezifischen Analyten zu liefern. Sie ist >90 %

10.4. Interferenzen

Hämolytische, lipämische und ikterische Proben ergaben bis zu einer Konzentration von 10 mg/ml für Hämoglobin, von 5 mg/ml Triglyceride und von 0,2 mg/ml für Bilirubin keine Interferenzen im vorliegenden ELISA.

Hinweis: Die Ergebnisse beziehen sich auf die untersuchten Probenkollektive; es handelt sich nicht um garantierte Spezifikationen.

11. GRENZEN DES VERFAHRENS

Die Familie der Flaviviridae beinhaltet die Serotypen des Dengue-Virus, das Gelbfiebervirus, das Japanische-B-Enzephalitis-Virus u. a. Kreuzreaktionen mit anderen Flaviviren sind möglich und sollten bei der Interpretation der Ergebnisse in Betracht gezogen werden. Es wird empfohlen positive Testergebnisse mit einer zweiten Testmethode zu bestätigen.

Kontamination der Proben durch Bakterien oder wiederholtes Einfrieren und Auftauen können zu einer Veränderung der Messwerte führen. Die Diagnose einer Infektionskrankheit darf nicht allein auf der Basis des Ergebnisses einer Bestimmung gestellt werden. Die anamnestischen Daten sowie die Symptomatologie des Patienten müssen zusätzlich zu den serologischen Ergebnissen in Betracht gezogen werden. Bei Immunsupprimierten und Neugeborenen besitzen die Ergebnisse der serologischen Tests nur einen begrenzten Wert.

12. SICHERHEITSMASSNAHMEN UND WARNHINWEISE

- Gemäß Art. 1 Abs. 2b der EU-Richtlinie 98/79/EG legt der Hersteller die Zweckbestimmung von In-vitro-Diagnostika fest, um deren Eignung, Leistung und Sicherheit sicherzustellen. Daher sind die Testdurchführung, die Information, die Sicherheitsmaßnahmen und Warnhinweise in der Gebrauchsanweisung strikt zu befolgen. Bei Anwendung des Testkits auf Diagnostika-Geräten ist die Testmethode zu validieren. Jede Änderung am Aussehen, der Zusammensetzung und der Testdurchführung sowie jede Verwendung in Kombination mit anderen Produkten, die der Hersteller nicht autorisiert hat, ist nicht zulässig; der Anwender ist für solche Änderungen selbst verantwortlich. Der Hersteller haftet für falsche Ergebnisse und Vorkommnisse aus solchen Gründen nicht. Auch für falsche Ergebnisse aufgrund von visueller Auswertung wird keine Haftung übernommen.
- Nur für in-vitro-Diagnostik.
- Alle verwendeten Bestandteile menschlichen Ursprungs sind auf Anti-HIV-AK, Anti-HCV-AK und HBsAG nicht-reakтив getestet. Dennoch sind alle Materialien als potentiell infektiös anzusehen und entsprechend zu behandeln.
- Reagenzien und Mikrotiterplatten unterschiedlicher Chargen nicht untereinander austauschen.
- Keine Reagenzien anderer Hersteller zusammen mit den Reagenzien dieses Testkits verwenden.
- Nicht nach Ablauf des Verfallsdatums verwenden.
- Nur saubere Pipettenspitzen, Dispenser und Labormaterialien verwenden.
- Verschlusskappen der einzelnen Reagenzien nicht untereinander vertauschen.
- Flaschen sofort nach Gebrauch fest verschließen, um Verdunstung und mikrobielle Kontamination zu vermeiden.
- Nach dem ersten Öffnen Konjugat- und Standardfläschchen vor weiterem Gebrauch auf mikrobielle Kontamination prüfen.
- Zur Vermeidung von Kreuzkontamination und falsch erhöhten Resultaten Patientenproben und Konjugat sorgfältig in die Kavitäten pipettieren.
- Der ELISA ist nur für die Anwendung durch Fachpersonal vorgesehen, welches die Arbeitstechniken einwandfrei beherrscht.

WARNUNG: Bronidox L zeigt in der verwendeten Konzentration nahezu keine toxikologischen Risiken an Haut bzw.

Schleimhaut.

WARNUNG: Schwefelsäure reizt Augen und Haut! Nach Berührung mit den Augen gründlich mit Wasser spülen und einen Arzt aufsuchen.

12.1. Entsorgungshinweise

Chemikalien und Zubereitungen sind in der Regel Sonderabfälle. Deren Beseitigung unterliegt den nationalen abfallrechtlichen Gesetzen und Verordnungen. Die zuständige Behörde informiert über die Entsorgung von Sonderabfällen.

BIBLIOGRAPHY / LITERATUR / BIBLIOGRAPHIE / BIBLIOGRAFIA / BIBLIOGRAFÍA

Hayes E.B.,Gubler D.J.,Dengue and dengue hemorrhagic fever,Pediatric Inf.Dis.J.11:311-317,1992

CDC: Imported Dengue-United States 1993-1994,J.Am.Med.Assn.274:113,1995

Rigau-Perez J.G.,Gubler D.J.,Vorndam A.V. et al,DEngue surveillance-United States 1986-1992, Morbidity and Mortality Weekly REport 43:7-19,1994

CDC: DEngue feveramong U.S. military personnel-Haiti,September-November,1994,J.Am.Med.Assn.,273:14-15,1994

Sharp T.W.,Wallace M.R.,Hayes C.G.et al,Dengue fever in US troops during Operation Restore Hope,Somalia 1992- 1993, Am.J.Trop.Med.Hyg.53:89-94,1995

SCHEME OF THE ASSAY

Dengue Virus IgG-ELISA

Test Preparation

Prepare reagents and samples as described.
 Establish the distribution and identification plan for all specimens and controls on the result sheet supplied in the kit.
 Select the required number of microtiter strips or wells and insert them into the holder.

Assay Procedure

	Substrate blank (e.g. A1)	Negative control	Positive control	Cut-off control	Sample (diluted 1+100)
Negative control	-	100µl	-	-	-
Positive control	-	-	100µl	-	-
Cut-off control	-	-	-	100µl	-
Sample (diluted 1+100)	-	-	-	-	100µl
Cover wells with foil supplied in the kit					
Incubate for 1 h at 37°C					
Wash each well three times with 300µl of washing solution					
Conjugate	-	100µl	100µl	100µl	100µl
Cover wells with foil supplied in the kit					
Incubate for 30 min at room temperature					
Wash each well three times with 300µl of washing solution					
TMB Substrate	100µl	100µl	100µl	100µl	100µl
Incubate for exactly 15 min at room temperature in the dark					
Stop Solution	100µl	100µl	100µl	100µl	100µl
Photometric measurement at 450 nm (reference wavelength: 620 nm)					

Symbols Key/ Symbolschlüssel/ Explication des symboles / Legenda / Símbolos	
	Manufactured by / Hergestellt von/ Fabriqué par/ Prodotto da/ Fabricado por
[IVD]	In Vitro Diagnostic Medical Device/ In Vitro Diagnosticum/ Dispositif médical de diagnostic <i>in vitro</i> / Diganostico <i>in vitro</i> / Producto para diagnóstico <i>In vitro</i>
LOT	Lot Number/ Chargenbezeichnung/ Numéro de lot/ Lotto/ Número de lote
	Expiration Date/ Verfallsdatum/ Date de péremption/ Scadenza/ Fecha de caducidad
	Storage Temperature/ Lagertemperatur/ Température de conservation/ Temperatura di conservazione / Temperatura de almacenamiento
CE	CE Mark/ CE-Zeichen/ Marquage CE / Marchio CE/ MarcaCE
[REF]	Catalogue Number/ Katalog Nummer/ Référence du catalogue/ Numero di codice/ Número de Catálogo
	Consult Instructions for Use/ Gebrauchsanweisung beachten/ Consulter la notice d'utilisation/ Consultare le istruzioni/ Consulte las Instrucciones de Uso
MTP	Microplate/ Mikrotiterplatte/ Microplaquette/ Micropiastra/ Microplaca
CONJ	Conjugate/ Konjugat/ Conjugué/ Coniugato/ Conjugado
CONTROL -	Control serum, negative/ Kontrollserum, negative/ Sérum de contrôle négatif/ siero di controllo, negativo /Suero control negativo/ Soro de control negativo
CONTROL +	Control serum, positive/ Kontrollserum, positiv/ Sérum de contrôle positif/ siero di controllo, positivo/ Suero de control positivo
CUT OFF	Cut off control serum/ Cut off Kontrollserum/ Sérum de contrôle du cut-off/ siero di controllo, cut-off/ Suero control Cut-off
DIL G	Sample diluent buffer IgG/ IgG-Probenverdünnungspuffer/ Tampon diluant pour échantillon IgG/ soluzione tampone per i campioni IgG/ solución tampón para muestras IgG
SOLN STOP	Stop solution/ Stoplösung/ Solution d'arrêt/Soluzione bloccante
SUB TMB	TMB Substrate solution/ TMB-Substratlösung/ Substrat TMB/ soluzione substrato TMB/ solución substrato TMB
WASHBUF 20x	Washing solution 20x concentrated/ Waschlösung 20x konzentriert/ Solution de lavage concentré 20 x/ soluzione di lavaggio concentrazione x20/ solución de lavado concentrado x20
	Contains sufficient for "n" tests/ Ausreichend für "n" Tests/ Contenu suffisant pour "n" tests/ Contenuto sufficiente per "n" saggi/ Contenido suficiente para "n" tests